



Declaration of Conformity CE, Reach and ROHS

Manufacturer:
Aven Inc.
4330 Varsity Drive
Ann Arbor, MI 48108

For the following equipment:

Item #	Description
26700-218, -PLR, -DFS	Mighty Scope V2 USB Digital Microscope, (with Polarizer), (with Diffuser)
26700-401	Cyclops Digital Microscope HDMI with 4x Lens
26700-423	Cyclops 3.0 Digital Microscope HDMI + USB with 4x Lens
26700-401-IMB-08	HDMI Image Measurement Box for Cyclops Digital HDMI
26700-220***	Mighty Scope ClearVue Digital Microscope***
26700-136	SharpVue Plus Digital Microscope
26700-137	SharpVue XT Digital Microscope
26700-140	MicroVue Digital Microscope with Built-In HD Monitor
26700-406	22" LCD Monitor

***= Includes all Mighty Scope ClearVue assembly options

- 26700-220-MNT, 26700-220-479, 26700-220-557, 26700-220-558, 26700-220-559, 26700-220-559WB, 26700-220-505, 26700-220-MNTD, 26700-220-479D, 26700-220-557D, 26700-220-558D, 26700-220-559D

This declaration of conformity is issued under the sole responsibility of the manufacturer.
The object of the declaration described above is in conformity with the relevant Union harmonization legislations and their amendments:

Low Voltage Directive 2014/35/EU
EMC Directive 2014/30/EU
RoHS Directive 2011/65/EU REACH
(EC) 1907/2006

References to the relevant harmonized standards used, including the date of the standard, or references to the other technical specifications, including the date of the specification, in relation to which conformity is declared:

Health & Safety	IEC 60335-2-29: 2016+A1:2019 IEC 60335-1: 2010+A1: 2013+A2:2016 IEC 61558-1:2005+A1 IEC 61558-2-16:2009+A1
EMC	EN 55014-1:2021 EN 61000-3-2: 2019+A1:2021 EN 61000-3-3:2013+A1:2019+A2:2021 EN 55014-2:2021
RoHS	IEC 63000:2018

Mike Shahpurwala

Mike Shahpurwala

Date: 3/21/2023

REVISION HISTORY

Rev.	Description	Date
1.0	Initial Release	3/21/2023

The revision history provided is for informational purposes only and is believed to be accurate.