

Manufacturer's Name: Liquid Instruments
Manufacturer's Address: 243 Northbourne Avenue
 Ground Floor, suite 2
 2602 Lyneham, ACT
 Canberra, Australia

Declares under sole responsibility that the product as originally delivered:

Product Name: Moku:Go
Model Number: Moku:GoV1.3
Product Option: Colours: Black, White, Storm, Fire, Sage, Blush
 Configurations: M0, M1, M2

complies with the essential requirements of the following applicable European Directives, and carries the UKCA marking accordingly:

2016 No. 1101	Electrical Equipment (Safety) Regulations 2016
2016 No. 1091	Electromagnetic Compatibility Regulations 2016
RoHS S.I. 2012 No. 3032	The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

And conforms with the following standard(s):

	Standard	Phenomena	Limit
EMC	IEC 61326-1:2013		
	CISPR 11		Group 1 Class A
	IEC 61000-4-2	Electrostatic discharge	4kV CD, 8kV AD
	IEC 61000-4-3	Electromagnetic field	10V/m 80MHz – 2GHz, 3V/m 1.4GHz - 27GHz
	IEC 61000-4-4	Burst	0.5kV signal lines, 1 kV power lines
	IEC 61000-4-5	Surge	0.5kV lineline, 1kV lineground, 1kV signal lines
	IEC 61000-4-6	Conducted RF	3V, 150kHz – 80 MHz
	IEC 61000-4-8	Power frequency magnetic field	3A/m; 50Hz, 60Hz
	IEC 61000-4-11	Voltage dips	100% for 0.5 (0°, 180°) cycle, 100% for 1 cycle, 30% for 25 cycles
	IEC 61000-4-11	Short interruptions	100% for 250 cycles

RoHS EN IEC 63000:2018

Supplementary Information:

- This product is intended for use in a basic environment.
- The products were tested in a typical configuration with Liquid Instruments test system.
- This is a sensitive measurement apparatus by design and may have some performance loss when exposed to ambient continuous electromagnetic phenomenon.

This DoC applies to the above listed products placed on the UK market after:

October 4, 2022

Signature



Date

Daniel Shaddock
CEO – Liquid Instruments