



Contents

Device	Part number	Quantity	Calibration Option ^a
Open circuit plug	03S12L-00FS3	1	FC
Open circuit jack	03K12L-00FS3	1	FC
Short circuit plug	03S12S-00FS3	1	FC
Short circuit jack	03K12S-00FS3	1	FC
Calibration load plug	03S150-F10S3	1	FC
Calibration load jack	03K150-F10S3	1	FC
Calibration adaptor RPC-3.50 jack / FAKRA-HF plug	03K159-S20S3	1	FC
Calibration adaptor RPC-3.50 jack / FAKRA-HF jack	03K159-K20S3	1	FC
Calibration adaptor RPC-3.50 plug / FAKRA-HF plug	03S159-S20S3	1	FC
Calibration adaptor RPC-3.50 plug / FAKRA-HF jack	03S159-K20S3	1	FC
Combi wrench	03W008-000	1	-
Torque wrench	03W021-000	1	FC
Plastic housing plug	59Z114-000Z	10	-
Plastic housing jack	59Z113-000Z	10	-

a. See "Declaration of calibration options" for explanation.

Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG

RF_35/05_10/6.1

Documentation

This kit is delivered with

- **USB-Stick**
Standard Definitions as data files for Vector Network Analyzer Families PNA (Keysight/Agilent) and ZVA (Rohde&Schwarz). Calibration Certificate as PDF-file.
- **Standard Definitions Cards**
Printed Standard Definitions that can be used on nearly all Vector Network Analyzers.
- **Kit Info Card**
Handling precautions and information for installing Standard Definitions on a Vector Network Analyzer.
- **Calibration Certificate**
Details see "Declaration of calibration options"
- **Operating Manual**

Electrical specifications

This specification covers electrical key values for the main calibration standards of the calibration kit. Specific datasheets are available for each component among the part number.

Calibration standard	Frequency	Parameter	Specification
Opens^b (plug and jack)	DC to ≤ 4 GHz > 4 GHz to ≤ 6 GHz	Error from Nominal Phase	≤ 1.0° ≤ 1.5°
Shorts^b (plug and jack)	DC to ≤ 4 GHz > 4 GHz to ≤ 6 GHz	Error from Nominal Phase	≤ 1.0° ≤ 1.5°
Calibration loads (plug and jack)	DC to ≤ 4 GHz > 4 GHz to ≤ 6 GHz	Return Loss	≥ 40 dB ≥ 35 dB
Calibration adaptors RPC-3.50 to FAKRA - HF (plug/plug ; plug/jack ; jack/plug ; jack/jack)	DC to ≤ 1 GHz > 1 GHz to ≤ 3 GHz > 3 GHz to ≤ 6 GHz	Return Loss	≥ 35 dB ≥ 26 dB ≥ 21 dB

b. The specifications for opens and shorts are given as allowed deviation from nominal model as defined in calibration certificate included with your kit.

Declaration of calibration options

Factory Calibration

Standard delivery for this kit includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, **traceable to Rosenberger standards**, national / international standards are not available. Model based standard definitions of the calibration standards are reported in Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

Accredited Calibration

Not available.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation 12 months

Recommended accessories

- Rosenberger Test Port Adaptor
- Rosenberger VNA Test cable kit and Microwave Cable Assemblies

For further, more detailed information please visit our homepage www.rosenberger.com.

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
H. Babinger	02.06.16	S. Andorfer	26.01.18	c00	17-1935	M. Knoll	26.01.18
Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de					Tel. : +49 8684 18-0 Fax : +49 8684 18-499 Email : info@rosenberger.de		Page 3 / 3

Mouser Electronics

Authorized Distributor

Click to View Pricing, Inventory, Delivery & Lifecycle Information:

[Rosenberger:](#)

[59CK010-150](#)