Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG

RF_35/09.14/6.2

Technical Data Sheet

Rosenberger

RPC-1.35

Calibration Kit Industrial Version

P9CK010-150



Contents

Device	Part number	Quantity	Calibration Option ^a	
Open circuit plug	P9S12L-000D3	1	FC	
Open circuit jack	P9K12L-000D3	1	FC	
Short circuit plug	P9S12S-000D3	1	FC	
Short circuit jack	P9K12S-000D3	1	FC	
Calibration load plug	P9S150-C10D3	1	FC	
Calibration load jack	P9K150-C10D3	1	FC	
Calibration adaptor plug/plug	P9S121-S20D3	1	FC	
Calibration adaptor jack/jack	P9K121-K20D3	1	FC	
Combi wrench	03W008-000	1	-	
Torque wrench	03W021-000	1	FC	

a. See "Declaration of calibration options" for explanation.

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.com

Tel. : +49 8684 18-0 Email : info@rosenberger.com Page

1/3

Technical Data Sheet		Rosenberger		
RPC-1.35	Calibration Kit Industrial Version	P9CK010-150		

Documentation

This kit is delivered with

USB-Stick

Standard Definitions as data files for Vector Network Anaylzer Families PNA (Keysight/Agilent) and ZVA (Rohde&Schwarz). S1P-files for Open, Short and Load calibration standards. Calibration Certificate as PDF-file.

Standard Definitions Cards

Model based Standard Definitions for the Calibration Adapters. Overview of electrical kit components.

Kit Info Card

Handling precautions and information for installing Standard Definitions on a Vector Network Analyzer.

• Calibration Certificate

Details see "Declaration of calibration options"

Operating Manual

Electrical specifications

These electrical specifications are only valid when the specific VNA files or the specific S1P-files are used as standard definitions. They include measurement uncertainties as well as guard bands to cover some tear and wear of the calibration standards.

Individual datasheets with further specifications and information are available for each component of this calibration kit.

Tel. : +49 8684 18-0

Email: info@rosenberger.com

Residual System Data*	Frequency	Specification (plug and jack)
Directivity	0.01 GHz to ≤ 40 GHz > 40 GHz to ≤ 80 GHz > 80 GHz to ≤ 90 GHz	≥ 33 dB ≥ 30 dB ≥ 27 dB
Source Match	0.01 GHz to ≤ 30 GHz > 30 GHz to ≤ 60 GHz > 60 GHz to ≤ 80 GHz > 80 GHz to ≤ 90 GHz	≥ 32 dB ≥ 28 dB ≥ 25 dB ≥ 23 dB
Reflection Tracking	0.01 GHz to ≤ 20 GHz > 20 GHz to ≤ 40 GHz > 40 GHz to ≤ 60 GHz > 60 GHz to ≤ 80 GHz > 80 GHz to ≤ 90 GHz	≤ 0.15 dB ≤ 0.20 dB ≤ 0.25 dB ≤ 0.35 dB ≤ 0.45 dB

^{*} Residual System Data are also called Effective System Data

Technical Data Sheet RPC-1.35 Calibration Kit Industrial Version RPC-1.35 P9CK010-150

Declaration of calibration options

Factory Calibration

Standard delivery for this kit includes a Factory Calibration. All devices marked with "FC" in the Content table above are reported in a Calibration Certificate with their individual calibration results, traceable to national / international standards. Data based definitions of the calibration standards are reported as data files for Vector Network Anaylzer Families PNA (Keysight/Agilent) and ZVA (Rohde&Schwarz) as well as S1P-files for Open, Short and Load calibration standards.

Accredited Calibration

Not available.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation

12 months

Recommended accessories

- Rosenberger Test Port Adaptor
- Rosenberger Gauge Kit P9GK0KS-010
- Rosenberger VNA Test cable kit and Microwave Cable Assemblies

For further, more detailed information please visit our homepage <u>www.rosenberger.com</u>.

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

For the installation of the electrotechnical equipment, particular electrotechnical expertise is required.



Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
F. Reiner	17.12.18	H. Babinger	17.04.23	d00	23-0667	T. Börgerding	17.04.23

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.com

Tel. : +49 8684 18-0 Email : info@rosenberger.com Page 3/3

Mouser Electronics

Authorized Distributor

Click to View Pricing, Inventory, Delivery & Lifecycle Information:

Rosenberger: P9CK010-150