


Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG  
RF\_35/09.14/6.2

|                                                                                                                                                    |               |                                                     |
|----------------------------------------------------------------------------------------------------------------------------------------------------|---------------|-----------------------------------------------------|
| Technical Data Sheet                                                                                                                               |               | Rosenberger                                         |
| RPC-1.00                                                                                                                                           | Torque Wrench | 01W021-000                                          |
| <div></div>                                                     |               |                                                     |
| All dimensions are in mm                                                                                                                           |               |                                                     |
| <div><div>Documents</div><div>Application noteAN001 “Calibration Services”</div></div>                                                             |               |                                                     |
| <div><div>Material and plating</div><div>Tooling partsSpannerMaterialStainless steel</div></div>                                                   |               |                                                     |
| <div><div>Mechanical data</div><div>Torque setting0.35 Nm ± 0.05 Nm (force applied in the middle of handle)<br/>Width across flats6 mm</div></div> |               |                                                     |
| Rosenberger Hochfrequenztechnik GmbH & Co. KG<br>P.O.Box 1260 D-84526 Tittmoning Germany<br>www.rosenberger.de                                     |               | Tel. : +49 8684 18-0<br>Email : info@rosenberger.de |
| öffentlich   public                                                                                                                                |               | Page<br>1 / 2                                       |

## Declaration of calibration options

## Factory Calibration

Standard delivery for this torque wrench includes a Factory Calibration. The Calibration Certificate issued reports individual mechanical calibration results, traceable to national / international standards.

## Accredited Calibration

Not available.

*For further, more detailed information see application note AN001 on the Rosenberger homepage.*

## Calibration interval

|                |           |
|----------------|-----------|
| Recommendation | 12 months |
|----------------|-----------|

## Packing

|          |              |
|----------|--------------|
| Standard | 1 pce in box |
| Weight   | 119.7 g/pce  |

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

|                                                                                                                                                        |          |                |          |                                                                                              |                           |         |               |
|--------------------------------------------------------------------------------------------------------------------------------------------------------|----------|----------------|----------|----------------------------------------------------------------------------------------------|---------------------------|---------|---------------|
| Draft                                                                                                                                                  | Date     | Approved       | Date     | Rev.                                                                                         | Engineering change number | Name    | Date          |
| Herbert Babinger                                                                                                                                       | 11.03.05 | Florian Reiner | 20.02.25 | f00                                                                                          | 25-0071                   | M. Gehl | 20.02.25      |
| Rosenberger Hochfrequenztechnik GmbH & Co. KG<br>P.O.Box 1260 D-84526 Tittmoning Germany<br><a href="http://www.rosenberger.de">www.rosenberger.de</a> |          |                |          | Tel. : +49 8684 18-0<br>Email : <a href="mailto:info@rosenberger.de">info@rosenberger.de</a> |                           |         | Page<br>2 / 2 |