




All dimensions are in mm

Documents

Application note

AN001 "Calibration Services"

Material and plating

Tooling parts

Spanner

Material

Stainless steel

Mechanical data

Torque setting

1.36 Nm $\pm$ 0.05 Nm (force applied in the middle of handle)

Width across flats

19 mm

Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG
RF_35/09.14/6.2

Technical Data Sheet				Rosenberger			
RPC-7		Torque Wrench		07W021-000			
Declaration of calibration options Factory Calibration Standard delivery for this torque wrench includes a Factory Calibration. The Calibration Certificate issued reports individual mechanical calibration results, traceable to national / international standards. Accredited Calibration Not available. For further, more detailed information see application note AN001 on the Rosenberger homepage. Calibration interval Recommendation12 months Packing Standard1 pce in box Weight153.7 g/pce							
While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.							
Draft		Date		Approved		Date	
Herbert Babinger		11.03.05		Florian Reiner		20.02.25	
Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de				Rev.		Engineering change number	
				f00		25-0071	
				Name		Date	
				M. Gehl		20.02.25	
				Tel. : +49 8684 18-0 Email : info@rosenberger.de			Page
							2 / 2